



Item number 1999902090

Classic


### Examination gloves, Abena Classic, XL, blue, nitrile, powder free

- ✓ Elastic
- ✓ Extra soft
- ✓ Adapts to the hand
- ✓ Thin for extra sensitivity
- ✓ blue



### Product Description

Elastic examination glove that fits to the shape of the hand during use and provides good sensitivity. Can be used for a variety of tasks such as cleaning, care, food handling, laboratory, medication handling, cytostatics and hormone creams. Contains no latex proteins therefore safe for use in the event of latex allergy.



### Specifications

Base name	Examination gloves
Brand	Abena
Sub-brand	Classic
Size	XL
Color	Blue
Properties	Finger textured, latex free, medium grip, rolled cuff, sensitive
Features	Powder free
Single or multiple use	Single use
Material	Nitrile
Technical ingredients	Nitrile NBR, Potassium Hydroxide, Zinc oxide, ZDEC, ZDBC, Sulfur, Titanium Dioxide, Wingstay L, Chlorine
Length/depth	240 mm
Width	115 mm
Thickness	0,08 (+/- 0,02 mm) mm
Certifications	CAT III, CE, Food contact materials, MD
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.5, EN 374-1:2016 Type B KPT, EN 374-5:2016 VIRUS, EN 420, EN 455
Directives, regulations and acts	MDR (EU) 2017/745, 425/2016/EU, 10/2011/EU
Because of the products composition it has an expiry date from the production date on:	5 years
Shelf Life, opened	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Sort as plastic. If contaminated dispose off as clinical waste.

## Instructions for use/application

Examine the gloves for flaws and defects before use.

## Packaging data

Unit	Contains	Length	Width	Height	EAN
pallet	72000 pcs	1.2 m	0.8 m	1.15 m	
cII	2000 pcs	0.38 m	0.256 m	0.25 m	5703538438103
pck	200 pcs	0.24 m	0.123 m	0.074 m	5703538438097
pair	2 pcs	0 m	0 m	0 m	



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. CAT III products will be accompanied by a notified body number for the 3rd party verification of the product's safety.



The glove has a medium grip.



The glove is 240 mm long.



The product does not contain latex.



The glove has a finger textured surface.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.

EN374-5:2016



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.

374-1:2016/  
Type B



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.

