

EU Declaration of Conformity

Legal Manufacturer

- Name	ABENA A/S	www.abena.com
- Address	Egelund 35	Phone: +45 7431 1818
- Phone / fax / email / webpage	DK-6200 Aabenraa	Fax: +45 7462 9737
	Denmark	Mail: info@abena.com
- Single Registration Number (SRN), if applicable	-	

Medical Device(s)

- Basic UDI-DI	-
- Product/trade name(s) and/or product code(s)	Nitril Medical Examination Gloves, Powder-Free
(REF)/and or catalogue number	
- Other ref. allowing identification (e.g. UDI-DI)	-
- Intended Purpose	Nitrile Examination gloves intended to be worn by
	health care practitioners to protect patient and user
	from cross-contamination.
- Risk classification	Class 1 according to rule no. 1 and 5 in MDR annex
	VIII

Other information (if applicable)

- Common Specifications used for compliance	
- Notified Body name and identification no. and	SATRA Technology Europe Limited
description of the conformity assessment	Bracetown Business Park
procedure performed	Clonee D15YN2P
	Ireland
- Additional information	EU Type Examination Module B and On-going
	conformity Module C2 performed by notified body
	2777, SATRA Technology Europe Limited.
	Issued EU Type Examination certificate number
	2777/10015-03/E06-01
- Standards used to assure compliance	EU 2016/425, EN374-1:2016, Type B, EN374-4:2013,
	EN374-5:2016 Virus, EN16523-1:2015

The above mention manufacturer hereby declare that the above mentioned medical device(s) are compliant with the EU Regulation 2017/745 for Medical Devices and the EU legislations mentioned under "additional information".

This Declaration of Conformity is issued under the sole responsibility of the above mentioned manufacturer.

Name and Function:	
Khalid Elamri, Global Category Manager	
Signature	Place and date of issue
Eldifferi	Aabenraa, DK, 15-01-2020
Abena, Egelund 35, DK-6200 Aabenraa	

Created: JOHO	Date: 23-01-2020
Copyright: Abena	Page 1/2



Appendix I, List of products

Product Name	Item no.	UDI-DI	Intended Purpose
Classic Sensitive Nitril Examination gloves powder free blue and white colour, 100 pcs	290610-290614 290471-290475	-	Protect patient and user from cross-contamination.

Appendix II, List of applicable standards used

Product Name	Standards used
Nitril Examination gloves powder free	EN 455-1:2002
Nitril Examination gloves powder free	EN 455-2:2015
Nitril Examination gloves powder free	EN 455-3:2015
Nitril Examination gloves powder free	EN 455-4:2009
Nitril Examination gloves powder free	EN374-1:2016
Nitril Examination gloves powder free	EN374-2:2003
Nitril Examination gloves powder free	EN374-4:2013
Nitril Examination gloves powder free	EN374-5:2016
Nitril Examination gloves powder free	EN16523-1:2015
Nitril Examination gloves powder free	EN420:2003

Created: JOHO	Date: 23-01-2020
Copyright: Abena	Page 2/2